

Company Name Firma Adı	AVAS KOZMETİK SANAYİ VE TİCARET LİMİTED ŞİRKETİ		
Company Address Firma Adresi	GÖKEVLER MAH. 2331 SK. PASİAD NO. 1 D ESENYURT - İSTANBUL / TÜRKİYE		
Test Name Testin Adı	Skin Sensitization Test		
Test Standard Test Standardı	TS EN ISO 10993-10:2014		
Commercial Brand (If You Have) Ticari Marka (Varsa)	Felix Filler		
Description of the Sample Numunenin Adı ve Tarifi	1 cc Felix Filler Dolgu		
Lot Number Lot Numarası	AVAS012023-1		
Sample Registration Number Numune Kayıt Numarası	FLXFHA/202302	Sample Acceptance Date Numune Kabul Tarihi	27.02.2023
Report Number Rapor Numarası	2023-03/BIYO/1532HA-DST	Date of Report Rapor Tarihi	10.03.2023
Date of Test Deney Tarihi	06.03.2023 - 09.03.2023		
Report Total Page Raporun Sayfa Sayısı	9 Page / Sayfa		

This test report has been prepared with the relevant report number as the result of the test made at the request of the customer. This test report is the number of pages mentioned above together with the result from the experiment device. The results belong to the sample with the above name and registration number. Since the test results are whole with output from the device and report given by MEDICERT, they cannot be quoted in sections from test results. This report has been prepared in duplicate to be given to a customer and one to stay in MEDICERT. The test report is signed by the laboratory staff and the laboratory officer. The device outputs are signed with initials of the responsible persons. The report without signature and paraphrase is invalid. The process of replicating the reports is the MEDICERT responsibility. The test report may not be reproduced in whole or in part. Medicert Test Laboratory is accredited by IAKS, a notified body based in the United States of America. The accreditation number is IAKS-TL-1018.

Bu deney raporu, yukarıda yazılı Numune Kabul Kayıt Numaralı talebe istinaden yapıları deney sonucunda ilgili Rapor Numarası ile hazırlanmıştır. Bu deney raporu, deney yapılan cihazdan alınan sonuç çıktısıyla birlikte yukarıda belirtilen sayfa sayısındadır. Sonuçlar yukarıda adı ve kayıt numarası yazan numuneye aittir. Deney sonuçları cihaz çıktısı (varsa) ve MEDICERT tarafından verilen bu raporla birlikte bir bütün olduğundan rapordan bölümler halinde alıntı yapılamaz. Bu rapor 1 adedi müşteriye verilmek üzere, 1 adedi de MEDICERT 'de kalmak üzere 2 adet hazırlanmıştır. Sonuç raporu deney yapan personel ve laboratuvar sorumlusu tarafından imzalanır. Cihaz çıktıları ise paraflanır, imza ve paraf bulunmayan rapor geçersizdir. Rapor çoğaltılması işlemi MEDICERT 'in sorumluluğundadır. Haricinde raporun tamamı veya bir kısmı çoğaltılamaz. Sonuç raporu MEDICERT 'in onayı alınmadan ticaret ve reklam amaçlı kullanılamaz. Medicert Test Laboratuvarı, Amerika Birleşik Devletleri merkezli onaylanmış kuruluş olan IAKS tarafından akredite edilmiştir. Akreditasyon numarası IAKS-TL-1018 'dir.

INDEX

SUMMARY	3
INTRODUCTION	4
SAMPLE INFORMATION	4
TEST SYSTEM	5
ANIMAL MANAGEMENT	5
METHOD	5
TEST GROUP	5
CONTROL GROUP	6
EVALUATION	7
RESULT	9
RECORD	9
REFERENCES	9

M E D I C E R T

SUMMARY

1 cc Felix Filler Dolgu sample number AVAS012023-1 has been subjected to biocompatibility testing according to TS EN ISO 10993-10 Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitivity. Samples were prepared by storing 37°C-72 hours in SF under sterile conditions according to TS EN ISO 10993-12 Biological evaluation of medical devices - Part 12: Sample preparation and reference materials standard. The extract was intradermally injected into 10 guinea pigs. 5 experimental animals were used as a control group and previously determined Serum Physiological was injected. On day 7, after the completion of the intradermal induction phase, the superficial induction phase was applied to both the test group and the control group. On the 14th day after the induction phase, the stimulation phase was applied to the test group and the control group. The dressings and patches were removed after 24 hours. Necessary evaluations were made according to Magnusson and Kligman ratings at 24th and 48th hours. Under the terms of this study, **1 cc Felix Filler Dolgu extract showed no evidence of sensitization.**

MEDICERT

1. INTRODUCTION

Purpose : This test was performed to evaluate the sensitivity of the sample described below to cause sensitivity.

Test Guide: This study was conducted according to the requirements of the International Organization for Standardization. 10993: Biological Assessment of Medical Devices, Part 10: Tests for Irritation and Skin Sensitivity

Dates

Sample Acceptance Date : 27.02.2023

Test Date : 06.03.2023

Observation Date : 06.03.2023 - 09.03.2023

2. SAMPLE INFORMATION

Company Name : Avas Kozmetik

Date of the Sample Acceptance : 27.02.2023 13.05

Sample Record Number : FLXFHA/202302

Sample Lot Number : AVAS012023-1

Number of Sample : 4

Packaging Information : CLOSED PACKED

Sample Delivery Method : CARGO


Expiration Date of the Sample : 17.01.2028

Production Date of the Sample : 17.01.2023

Description of the Sample : Bir şırınga 1 cc Felix Filler dolgu maddesi içermektedir.

Characteristics of the Sample : Cross linked hyaluronic acid
Use/Application :

Sample Image :



3. TEST SYSTEM

Animal used in the test : Guinea Pig
Strain : Dunkin Hartley
Source : Burdur Mehmet Akif Ersoy University Experimental
Animals Production and Research Center
Gender : Male
Weight : 400 - 500 Gr
Age : 10 - 12 Week
Acclimation time : 5 Days
Number of the animals : 15

4. ANIMAL MANAGEMENT

Animal Care : The animals used in the experiments are performed in accordance with the standards of Biological Evaluation of Medical Devices - Part 2: Requirements for Animal Welfare.

Food : The SDS brand VRF1 diet is provided as ad-libitum.

Water : Water is supplied as ad-libitum in suitable drinkers.

Cage System : Each animal was identified and placed in appropriate cages.

Environmental Conditions : 12 hours night and 12 hours day environment is provided; 30-70% humidity and 16-22°C environment is provided. Temperature and humidity are checked daily.

Personnel : Tests are performed by trained and appropriately qualified personnel.

Selection of the animal : Healthy, non-disease animals and non-pregnant animals were used under the supervision of a veterinary surgeon.

Veterinary Care : This study was carried out under the supervision of a veterinarian.

5. METHOD

Sensitization tests; The Intradermal Inducing Phase was completed by observing the animals and evaluating the results after the Superficial Inducing Phase and Stimulation Phase.

In the Intradermal Inducing Phase; 0.1 ml intradermal applications were made with the following ingredients.

TEST GROUP:

Figure 1 Area A: Mixture of NOVAFIX ELASTIC FIXING TAPE A3294 -10 HX10 CM and FCA (Freund's Complete Adjuvant) at 50:50 volume ratio

Figure 1 Area B: Test Sample (Undiluted Extract) in Test Group

Figure 1 Area C: Mixtures of materials applied in Areas A and B in a 50:50 volume ratio in the Test Group

CONTROL GROUP:

Figure 2 Area A: -

Figure 2 Area B: SERUM PHYSIOLOGICAL

Figure 2 Area C: Mixture of SERUM PHYSIOLOGY and FCA (Freund's Complete Adjuvant) at 50:50 volume ratio

Superficial Inducing Phase; On the 7th day after the completion of the intradermal induction phase; To the TEST GROUP; Superficial application is made to the infrascapular region with 8 cm square absorbent gauze to cover the intradermal injection sites. In the Transcutaneous Inducing Phase, the concentration determined in the Test Group Area B is used.

to the CONTROL GROUP; only SERUM PHYSIOLOGICAL is applied.

The dressing and patches are removed after 48 hours.

Stimulation Phase; After the completion of the superficial induction phase, on the 14th day, to the TEST GROUP; Superficial application is made to the untreated areas of the experimental animals and covered with absorbent gauze. In the Intradermal Inducing Phase, the concentration determined in the Test Group Area C is used.

to the CONTROL GROUP; only SERUM PHYSIOLOGICAL is applied. The dressing and patches are removed after 24 hours.

At 24 hours and 48 hours after the completion of the stimulation phase, stimulation sites are observed on the skin of the animals in the test and control groups. Observations are carried out under full spectrum lighting. Skin reactions for Erythema and Edema are completed and graded according to the Magnusson and Kligman grading specified in Table 1 for each time slot for each stimulation site.




Table 1. Magnusson and Kligman rating scale

Patch Test Reaction	Rating Scale
No visible changes	0
Discrete or patchy erythema	1
Moderate or adjacent erythema	2
Evident	3
Severe erythema or swelling	4

6. EVALUATION

Table 2. Results of the test and control groups at 24 and 48 hours depending on the Magnusson and Kligman rating scale

G R O U P	GUINEA PIG NO	AREA	MAGNUSSON AND KLIMMAN RATING ERYTHEMA / PAYMENT	
			EVALUATION TIME POINT	
			24th h	48th h
1 c c F e l i x F i l l e r D o l g u	1	A	2	1
		B	0	0
		C	1	0
	2	A	2	1
		B	0	0
		C	1	1
	3	A	2	1
		B	0	0
		C	1	1
	4	A	2	2
		B	0	0
		C	1	1
	5	A	2	2
		B	0	0
		C	1	1
	6	A	2	1
		B	0	0
		C	1	1
	7	A	2	2
		B	0	0
		C	2	1
	8	A	1	1
		B	0	0
		C	1	0
	9	A	2	1
		B	0	0
		C	1	1
	10	A	2	1
		B	0	0
		C	1	1

G R O U P	GUINEA PIG NO	AREA	MAGNUSSON AND KLIMMAN RATING ERYTHEMA / PAYMENT	
			EVALUATION TIME POINT	
			24th h	48th h
C O N T R O L G R O U P	1	A	-	-
		B	0	0
		C	1	1
	2	A	-	-
		B	0	0
		C	1	1
	3	A	-	-
		B	0	0
		C	1	1
	4	A	-	-
		B	0	0
		C	1	1
	5	A	-	-
		B	0	0
		C	1	1

Table 3. Mean Scoring of the values of Table 2

GROUPS	Mean Results
Test Group	0.82
Control Group	0.48

Table 4. End of test weight table of the test animals

TEST GROUP Guinea pig numbers	1	2	3	4	5	6	7	8	9	10
Weights at the end of the test	425	432	437	466	451	431	459	434	499	467
CONTROL GROUP Guinea pig numbers	1		2		3		4		5	
Weights at the end of the test	460		448		435		449		450	

7. RESULT

Following the tests, observations were made in two different time periods as stated. The mean score was obtained by averaging the values obtained. The sensitization score was 0.87 for the sample samples tested. It was determined that the health conditions of the animals were not good and there was no significant weight loss. According to the obtained results, **it was determined that the tested samples do not have sensitizing properties** based on the protocol and evaluation criteria specified in ISO 10993-10.

8. RECORD

All raw data and a copy of the final report are stored in the Medicert archive files.

9. REFERENCES

- ❖ Guide for The Care and Use of Laboratory Animals Eighth Edition National Research Council of The National Academies
- ❖ TS EN ISO 10993-1 Biological evaluation of Medical Devices - Chapter 1: Evaluation and experiment in a risk management process
- ❖ TS EN ISO 10993-2 Biological evaluation of Medical Devices Chapter 2: Conditions for animal welfare
- ❖ TS EN ISO 10993-10 Biological evaluation of Medical Devices – Chapter 10: Experiments for irritation and skin sensitivity
- ❖ TS EN ISO 10993-12 Biological evaluation of Medical Devices Chapter 12: Sample preparation and reference materials

Responsible for Laboratory
Laboratuvar Sorumlusu

Laboratory Manager
Laboratuvar Müdürü

Sample Acceptance Officer
Numune Kabul Sorumlusu

Veterinarian Simge GARLI

Erol ÜSTÜN

Gonca AÇILMIŞ



MEDICERT ULUSLARARASI ÜRÜN VE SİSTEM
BELGELENDİRME BAĞIMSIZ DENETİM
VE EĞİTİM HİZMETLERİ LTD. ŞTİ.
Tersane Mah. Cemal Gürsel Cad. No: 11/3
Halide Hanım Apt. Karşıyaka / İZMİR
Tel: 0232 327 33 44 Fax: 0232 327 33 45
Karşıyaka V.D: 613 073 9815



10.03.2023